

BioVendor

Research
and Diagnostic Products



CANINE CLUSTERIN ELISA

Product Data Sheet

Cat. No.: RD491034200R

For Research Use Only

CONTENTS

1.	INTENDED USE	3
2.	STORAGE, EXPIRATION	3
3.	INTRODUCTION	4
4.	TEST PRINCIPLE	5
5.	PRECAUTIONS	5
6.	TECHNICAL HINTS	6
7.	REAGENT SUPPLIED	6
8.	MATERIAL REQUIRED BUT NOT SUPPLIED	7
9.	PREPARATION OF REAGENTS	7
10.	PREPARATION OF SAMPLES	9
11.	ASSAY PROCEDURE	10
12.	CALCULATIONS	12
13.	PERFORMANCE CHARACTERISTICS	13
14.	DEFINITION OF THE STANDARD	15
15.	METHOD COMPARISON	15
16.	TROUBLESHOOTING AND FAQs	16
17.	REFERENCES	17
18.	EXPLANATION OF SYMBOLS	19

**➤➤ This kit is manufactured by:
BioVendor – Laboratorní medicína a.s.**

➤➤ Use only the current version of Product Data Sheet enclosed with the kit!

1. INTENDED USE

The RD491034200R Canine Clusterin ELISA is a sandwich enzyme immunoassay for the quantitative measurement of canine clusterin.

»» Features

- **It is intended for research use only**
- The total assay time is less than 3.5 hours
- The kit measures clusterin in canine serum and urine
- Assay format is 96 wells
- Quality Controls are canine serum based. No human sera are used
- Standard is canine serum based native protein
- Components of the kit are provided ready to use, concentrated or lyophilized

2. STORAGE, EXPIRATION

Store the complete kit at 2-8°C. Under these conditions, the kit is stable until the expiration date (see label on the box).

For stability of opened reagents see Chapter 9.

3. INTRODUCTION

Clusterin (Apolipoprotein J; SP-40,40; TRPM-2; SGP-2; pADHC-9; CLJ; T64; GP III; XIP8) is a highly conserved disulfide-linked secreted heterodimeric glycoprotein of 75-80 kDa but truncated forms targeted to nucleus have also been identified.

The protein is constitutively secreted by a number of cell types including epithelial and neuronal cells and is a major protein in physiological fluids including plasma, milk, urine, cerebrospinal fluid and semen. Due to its wide tissue distribution many diverse physiological functions have been attributed to clusterin including sperm maturation, membrane recycling, lipid transportation, tissue remodelling, complement inhibition and cell-cell or cell-substratum interactions. Moreover, it was proposed, that clusterin functions as an extra cellular chaperon that stabilizes stressed proteins in a folding-competent state and protein has also been implicated in programmed cell death. Another defining prominent of clusterin is its induction in many severe physiological disturbances states including kidney degenerative diseases, prostate and vesicle carcinogenesis, ovarian cancer and several neurodegenerative conditions (Alzheimer's disease).

Recent study demonstrate, that serum clusterin level increases significantly in diabetic type II patients and in patients with developing coronary heart disease, or myocardial infarction. These data raise the possibility that elevated clusterin levels in serum may represent a strong indication of vascular damage.

In patients with systemic lupus erythematosus (SLE) was found reduced serum clusterin levels that correlated inversely with disease activity. Lowered clusterin levels could be involved in the pathogenesis of SLE on account of decreased protective effects.

Another interesting observations obtain in rat model suggest that measurement of urinary clusterin levels may be a useful clinical valuable marker for the severity of renal tubular damage. Furthermore, urinary clusterin may also help to differentiate between tubular and glomerular forms of proteinuria.

Areas of investigation:

Coronary heart diseases

Myocardial infarction

Neurodegenerative diseases

Kidney degenerative disease

Renal tubular damage

4. TEST PRINCIPLE

In the BioVendor Canine Clusterin ELISA, standards, quality controls and samples are incubated in microplate wells pre-coated with polyclonal anti-canine clusterin antibody. After 60 minutes incubation and washing, biotin labelled polyclonal anti-canine clusterin antibody is added and incubated with captured clusterin for 60 minutes. After another washing, streptavidin-HRP conjugate is added. After 30 minutes incubation and the last washing step, the remaining conjugate is allowed to react with the substrate solution (TMB). The reaction is stopped by addition of acidic solution and absorbance of the resulting yellow product is measured. The absorbance is proportional to the concentration of clusterin. A standard curve is constructed by plotting absorbance values against concentrations of standards, and concentrations of unknown samples are determined using this standard curve.

5. PRECAUTIONS

- **For professional use only**
- Wear gloves and laboratory coats when handling immunodiagnostic materials
- Do not drink, eat or smoke in the areas where immunodiagnostic materials are being handled
- This kit contains components of animal origin. These materials should be handled as potentially infectious
- Avoid contact with the acidic Stop Solution and Substrate Solution, which contains hydrogen peroxide and tetramethylbenzidine (TMB). Wear gloves and eye and clothing protection when handling these reagents. Stop and/or Substrate Solutions may cause skin/eyes irritation. In case of contact with the Stop Solution and the Substrate Solution wash skin/eyes thoroughly with water and seek medical attention, when necessary
- The materials must not be pipetted by mouth

6. TECHNICAL HINTS

- Reagents with different lot numbers should not be mixed
- Use thoroughly clean glassware
- Use deionized (distilled) water, stored in clean containers
- Avoid any contamination among samples and reagents. For this purpose, disposable tips should be used for each sample and reagent
- Substrate Solution should remain colourless until added to the plate. Keep Substrate Solution protected from light
- Stop Solution should remain colourless until added to the plate. The colour developed in the wells will turn from blue to yellow immediately after the addition of the Stop Solution. Wells that are green in colour indicate that the Stop Solution has not mixed thoroughly with the Substrate Solution
- Dispose of consumable materials and unused contents in accordance with applicable national regulatory requirements

7. REAGENT SUPPLIED

<i>Kit Components</i>	<i>State</i>	<i>Quantity</i>
Antibody Coated Microtiter Strips	ready to use	96 wells
Biotin Labelled Antibody Conc. (50x)	concentrated	0.28 ml
Streptavidin-HRP Conjugate	ready to use	13 ml
Master Standard	lyophilized	2 vials
Quality Control HIGH	lyophilized	2 vials
Quality Control LOW	lyophilized	2 vials
Biotin-Ab Diluent	ready to use	13 ml
Dilution Buffer	ready to use	50 ml
Wash Solution Conc. (10x)	concentrated	100 ml
Substrate Solution	ready to use	13 ml
Stop Solution	ready to use	13 ml
Product Data Sheet + Certificate of Analysis	-	1 pc

8. MATERIAL REQUIRED BUT NOT SUPPLIED

- Deionized (distilled) water
- Test tubes for diluting samples
- Glassware (graduated cylinder and bottle) for Wash Solution (Dilution Buffer)
- Precision pipettes to deliver 5-1000 μl with disposable tips
- Multichannel pipette to deliver 100 μl with disposable tips
- Absorbent material (e.g. paper towels) for blotting the microtiter plate after washing
- Vortex mixer
- Orbital microplate shaker capable of approximately 300 rpm
- Microplate washer (optional). [Manual washing is possible but not preferable.]
- Microplate reader with 450 ± 10 nm filter, preferably with reference wavelength 630 nm (alternatively another one from the interval 550-650 nm)
- Software package facilitating data generation and analysis (optional)

9. PREPARATION OF REAGENTS

- All reagents need to be brought to room temperature prior to use
 - Always prepare only the appropriate quantity of reagents for your test
 - Do not use components after the expiration date marked on their label
- Assay reagents supplied ready to use:

Antibody Coated Microtiter Strips

Stability and storage:

Return the unused strips to the provided aluminium zip-sealed bag with desiccant and seal carefully. Remaining Microtiter Strips are stable 3 months stored at 2-8°C and protected from the moisture.

Streptavidin-HRP Conjugate

Biotin-Ab Diluent

Dilution Buffer

Substrate Solution

Stop Solution

Stability and storage:

Opened reagents are stable 3 months when stored at 2-8°C.

- **Assay reagents supplied concentrated or lyophilized:**

Canine Clusterin Master Standard

Refer to the Certificate of Analysis for current volume of Dilution Buffer needed for reconstitution of standard!!!

Reconstitute the lyophilized Master Standard with Dilution Buffer just prior to the assay. Let it dissolve at least 15 minutes with occasional gentle shaking (not to foam). The resulting concentration of the clusterin in the stock solution is **128 ng/ml**.

Prepare set of standards using Dilution Buffer as follows:

<i>Volume of Standard</i>	<i>Dilution Buffer</i>	<i>Concentration</i>
Stock	-	128 ng/ml
250 µl of stock	250 µl	64 ng/ml
250 µl of 64 ng/ml	250 µl	32 ng/ml
250 µl of 32 ng/ml	250 µl	16 ng/ml
250 µl of 16 ng/ml	250 µl	8 ng/ml
250 µl of 8 ng/ml	250 µl	4 ng/ml

Prepared Standards are ready to use, do not dilute them.

Stability and storage:

Do not store the Standard stock solution and set of standards.

Quality Controls HIGH, LOW

Refer to the Certificate of Analysis for current volume of Dilution Buffer needed for reconstitution and for current Quality Control concentration!!!

Reconstitute each Quality Control (HIGH and LOW) with Dilution Buffer just prior to the assay. Let it dissolve at least 15 minutes with occasional gentle shaking (not to foam).

Reconstituted Quality Controls are ready to use, do not dilute them.

Stability and storage:

Do not store the reconstituted Quality Controls.

Note:

Concentration of analyte in Quality Controls need not be anyhow associated with normal and/or pathological concentrations in serum or another body fluid. Quality Controls serve just for control that the kit works in accordance with PDS and CoA and that ELISA test was carried out properly.

Biotin Labelled Antibody Conc. (50x)

Prepare the working Biotin Labelled Antibody solution by adding 1 part Biotin Labelled Antibody Concentrate (50x) with 49 parts Biotin-Ab Diluent. Example: 20 µl of Biotin Labelled Antibody Concentrate (50x) + 980 µl of Biotin-Ab Diluent for 1 strip (8 wells).

Stability and storage:

Opened Biotin Labelled Antibody Concentrate (50x) is stable 3 months when stored at 2-8°C.

Do not store the diluted Biotin Labelled Antibody solution.

Wash Solution Conc. (10x)

Dilute Wash Solution Concentrate (10x) ten-fold in distilled water to prepare a 1x working solution. Example: 100 ml of Wash Solution Concentrate (10x) + 900 ml of distilled water for use of all 96-wells.

Stability and storage:

The diluted Wash Solution is stable 1 month when stored at 2-8°C. Opened Wash Solution Concentrate (10x) is stable 3 months when stored at 2-8°C.

10. PREPARATION OF SAMPLES

The kit measures clusterin in canine serum and urine.

Samples should be assayed immediately after collection or should be stored at -20°C. Mix thoroughly thawed samples just prior to the assay and avoid repeated freeze/thaw cycles, which may cause erroneous results. Avoid using hemolyzed or lipemic samples.

Serum samples:

Dilute samples just prior to perform the test 2 500x with Dilution Buffer in two steps as follows:

Dilution A (50x):

Add 5 µl of sample into 245 µl of Dilution Buffer and **mix well** (not to foam). Vortex is recommended.

Dilution B (50x):

Add 5 µl of Dilution A into 245 µl of Dilution Buffer to prepare final dilution 2 500x. **Mix well** (not to foam). Vortex is recommended.

Urine samples:

Dilute urine samples just prior to perform the test 3x with Dilution Buffer, e.g. 50 µl of sample + 100 µl of Dilution Buffer for singlets or 80 µl of sample + 160 µl of Dilution Buffer for duplicates. **Mix well** (not to foam). Vortex is recommended.

Results exceeding urine clusterin level of 64 ng/ml should be repeated with more diluted samples. It is recommended to measure this urine sample in the same assay in two dilutions 30x and 300x, e.g.:

10 µl of sample + 290 µl of Dilution Buffer for final dilution 30x

25 µl of dilution 30x + 225 µl of Dilution Buffer for final dilution 300x

Dilution factor have to be taken into consideration in calculating the clusterin concentration.

Stability and storage:

Serum samples should be stored at -20°C, or preferably at -70°C for long-term storage. Urine samples should be stored at -70°C. Avoid repeated freeze/ thaw cycles.

Do not store the diluted samples.

Note: It is recommended to use a precision pipette and a careful technique to perform the dilution in order to get precise results.

11. ASSAY PROCEDURE

1. Pipet **100 µl** of diluted Standards, Quality Controls, Dilution Buffer (=Blank) and samples, preferably in duplicates, into the appropriate wells. See *Figure 1* for example of work sheet.
2. Incubate the plate at room temperature (ca. 25°C) for **1 hour**, shaking at ca. 300 rpm on an orbital microplate shaker.
3. Wash the wells 3-times with Wash Solution (0.35 ml per well). After final wash, invert and tap the plate strongly against paper towel.
4. Add **100 µl** of Biotin Labelled Antibody solution into each well.
5. Incubate the plate at room temperature (ca. 25°C) for **1 hour**, shaking at ca. 300 rpm on an orbital microplate shaker.
6. Wash the wells 3-times with Wash Solution (0.35 ml per well). After final wash, invert and tap the plate strongly against paper towel.
7. Add **100 µl** of Streptavidin-HRP Conjugate into each well.
8. Incubate the plate at room temperature (ca. 25°C) for **30 min**, shaking at ca. 300 rpm on an orbital microplate shaker.
9. Wash the wells 3-times with Wash Solution (0.35 ml per well). After final wash, invert and tap the plate strongly against paper towel.
10. Add **100 µl** of Substrate Solution into each well. Avoid exposing the microtiter plate to direct sunlight. Covering the plate with e.g. aluminium foil is recommended.
11. Incubate the plate for **10 minutes** at room temperature. The incubation time may be extended [up to 20 minutes] if the reaction temperature is below than 20°C. Do not shake the plate during the incubation.
12. Stop the colour development by adding **100 µl** of Stop Solution.
13. Determine the absorbance of each well using a microplate reader set to 450 nm, preferably with the reference wavelength set to 630 nm (acceptable range: 550 – 650 nm). Subtract readings at 630 nm (550 - 650 nm) from the readings at 450 nm. **The absorbance should be read within 5 minutes following step 12.**

Note: If some samples and standard/s have absorbances above the upper limit of your microplate reader, perform a second reading at 405 nm. A new standard curve, constructed using the values measured at 405 nm, is used to determine clusterin concentration of off-scale standards and samples. The readings at 405 nm should not replace the readings for samples that were “in range” at 450 nm.

Note 2: Manual washing: Aspirate wells and pipet 0.35 ml Wash Solution into each well. Aspirate wells and repeat twice. After final wash, invert and tap the plate strongly against paper towel. Make certain that Wash Solution has been removed entirely.

	strip 1+2	strip 3+4	strip 5+6	strip 7+8	strip 9+10	strip 11+12
A	Standard 128	Blank	Sample 8	Sample 16	Sample 24	Sample 32
B	Standard 64	Sample 1	Sample 9	Sample 17	Sample 25	Sample 33
C	Standard 32	Sample 2	Sample 10	Sample 18	Sample 26	Sample 34
D	Standard 16	Sample 3	Sample 11	Sample 19	Sample 27	Sample 35
E	Standard 8	Sample 4	Sample 12	Sample 20	Sample 28	Sample 36
F	Standard 4	Sample 5	Sample 13	Sample 21	Sample 29	Sample 37
G	QC HIGH	Sample 6	Sample 14	Sample 22	Sample 30	Sample 38
H	QC LOW	Sample 7	Sample 15	Sample 23	Sample 31	Sample 39

Figure 1: Example of a work sheet.

12. CALCULATIONS

Most microplate readers perform automatic calculations of analyte concentration. The standard curve is constructed by plotting the mean absorbance (Y) of Standards against the known concentration (X) of Standards in logarithmic scale, using the four-parameter algorithm. Results are reported as concentration of clusterin ng/ml in samples.

Alternatively, the *logit log* function can be used to linearize the standard curve, i.e. *logit* of the mean absorbance (Y) is plotted against log of the known concentration (X) of Standards.

The measured concentration of samples calculated from the standard curve must be multiplied by their respective dilution factor, because samples have been diluted prior to the assay, e.g. 20 ng/ml (from standard curve) x 2 500 (dilution factor) = 50 000 ng/ml = 50 µg/ml.




Figure 2: Typical Standard Curve for Canine Clusterin ELISA.

13. PERFORMANCE CHARACTERISTICS

➤➤ Typical analytical data of BioVendor Canine Clusterin ELISA are presented in this chapter

- **Sensitivity**

Limit of Detection (LOD) (defined as concentration of analyte giving absorbance higher than mean absorbance of blank* plus three standard deviations of the absorbance of blank: $A_{\text{blank}} + 3 \times \text{SD}_{\text{blank}}$) is calculated from the real canine clusterin values in wells and is 0.2 ng/ml.

*Dilution Buffer is pipetted into blank wells.

- **Limit of assay**

Results exceeding clusterin level of 128 ng/ml should be repeated with more diluted samples. Dilution factor needs to be taken into consideration in calculating the clusterin concentration.

Results from assaying of urine samples exceeding clusterin level of 64 ng/ml should be repeated with more diluted urine samples for correct values. Dilution factor needs to be taken into consideration in calculating the clusterin concentration.

- **Specificity**

The antibodies used in this ELISA are specific for canine clusterin.

Sera of several mammalian species were measured in the assay. See results below.

For details please contact us at info@biovendor.com.

<i>Mammalian serum sample</i>	<i>Observed crossreactivity</i>
Bovine	no
Cat	yes
Goat	no
Hamster	no
Horse	no
Human	no
Monkey	no
Mouse	no
Pig	no
Rabbit	no
Rat	no
Sheep	no

➤➤ **Presented results are multiplied by respective dilution factor**

• **Precision**

Intra-assay (Within-Run) (n=8)

<i>Sample</i>	<i>Mean ($\mu\text{g/ml}$)</i>	<i>SD ($\mu\text{g/ml}$)</i>	<i>CV (%)</i>
1	50.5	2.6	5.1
2	62.5	2.1	3.4

Inter-assay (Run-to-Run) (n=8)

<i>Sample</i>	<i>Mean ($\mu\text{g/ml}$)</i>	<i>SD ($\mu\text{g/ml}$)</i>	<i>CV (%)</i>
1	28.8	2.1	7.4
2	37.5	1.9	5.0

• **Spiking Recovery**

Serum samples were spiked with different amounts of clusterin and assayed.

<i>Sample</i>	<i>Observed ($\mu\text{g/ml}$)</i>	<i>Expected ($\mu\text{g/ml}$)</i>	<i>Recovery O/E (%)</i>
1	55.6	-	-
	129.2	135.6	95.3
	87.3	95.6	91.4
	70.6	75.6	93.5
2	63.1	-	-
	130.2	143.1	91.0
	94.4	103.1	91.5
	79.5	83.1	95.6

• **Linearity**

Serum samples were serially diluted with Dilution Buffer and assayed.

<i>Sample</i>	<i>Dilution</i>	<i>Observed ($\mu\text{g/ml}$)</i>	<i>Expected ($\mu\text{g/ml}$)</i>	<i>Recovery O/E (%)</i>
1	-	43.0	-	-
	2x	22.1	21.5	102.7
	4x	10.5	10.8	98.0
	8x	5.0	5.4	93.8
2	-	54.0	-	-
	2x	29.0	27.0	107.3
	4x	14.1	13.5	104.3
	8x	7.4	6.8	109.0

- **Reference range**

It is recommended that each laboratory include its own panel of control samples in the assay. Each laboratory should establish its own normal and pathological reference ranges for canine clusterin levels with the assay.

14. DEFINITION OF THE STANDARD



Standard in this assay is canine serum based native clusterin.

15. METHOD COMPARISON



BioVendor Canine Clusterin ELISA has not been compared to any other immunoassay.

16. TROUBLESHOOTING AND FAQs

»» Weak signal in all wells

Possible explanations:

- Omission of a reagent or a step
- Improper preparation or storage of a reagent
- Assay performed before reagents were allowed to come to room temperature
- Improper wavelength when reading absorbance

»» High signal and background in all wells

Possible explanations:

- Improper or inadequate washing
- Overdeveloping; incubation time with Substrate Solution should be decreased before addition of Stop Solution
- Incubation temperature over 30°C

»» High coefficient of variation (CV)

Possible explanation:

- Improper or inadequate washing
- Improper mixing Standards, Quality Controls or samples

17. REFERENCES









»» References to clusterin:

- Choi-Miura NH, Oda T: Relationship between multifunctional protein Clusterin and Alzheimer disease. *Neurobiol. Aging* 1996; 17(5): 717-722
- Newkirk MM, Apostolakos P, Neville C and Fortin PR: Systemic lupus erythematosus, a disease associated with low levels of Clusterin/ApoJ, and anti-inflammatory protein. *J. Rheumatol.* 1999; 3:597-603
- Morrissey C, Lakins J, Moquin A, Hussain M, Tenniswood M: An antigen capture assay for the measurement of serum Clusterin concentrations. *J. Biochen. Biophys. Methods* 2001; 48:13-21
- Girton RA, Sundin DP, and Rosenberg ME: Clusterin protects renal tubular epithelial cells from gentamicin-mediated cytotoxicity. *Am J Renal Physiol.* 2002; 282:F709-F709
- Trougakos IP, Poulakou M, Stathatos M, Chalikia A, Melidonis A, Gonos ES: Serum levels of the senescence biomarker Clusterin/apolipoprotein J increase significantly in diabetes type II and during development of coronary heart disease or at myocardial infarction. *Ex. Gerontology* 2002; 37: 1175-1187
- Jones SE, Jomary C: Molecules in focus Clusterin. *The International J. of Bioch. & Cell Biol.* 2002; 34:427-431
- Hidaka S, Kränzlin B, Gretz N, Witzgall R: Urinary Clusterin levels in the rat correlate with the severity of tubular damage and may help to differentiate between glomerular and tubular injuries. *Cell Tissue Res.* 2002; 310:289-296
- Ghiggeri GM, Bruschi M, Candiano G, Rastaldi MP, Scolari F, Passerini P, Musante L, Pertica N, Caridi G, Ferrario F, Perfumo F, and Ponticelli C: Depletion of clusterin in renal diseases causing nephrotic syndrome. *Kidney Intern.* 2002; 62:2184-2194
- Chen X, Halberg RB, Ehrhardt WM, Torrealba J and Dove WF: Clusterin as a biomarker in murine and human intestinal neoplasia. *PNAS* 2003; 100:9530-9535
- Zhang LY, Ying WT, Mao YS, He HZ, Liu Y, Wang HX, Liu F, Wang K, Zhang DC, Wang Y, Wu M, Qian XH and Zhao XH: Loss of Clusterin both in serum and tissue correlates with the tumorigenesis of esophageal squamous cell carcinoma via proteomics approaches. *World J Gastroenterol* 2003; 9:650-654
- Wang L, Erling P, Bengtsson AA, Truedsson L, Sturfelt G, Erlinge D: Transcriptional down-regulation of the platelet ADP receptor P2Y₁₂ and Clusterin in patients with systemic lupus erythematosus. *J. of Thromb. And Haemost.* 2004; 2:1436-1442
- Patel NV, Wei M, Wong A, Finch CE, Morgan TE: Progressive changes in regulation of apolipoproteins E and J in glial cultures during postnatal development and aging. *Neuroscience Letters* 2004; 371:199-204
- Krijnen PAJ, Cillessen SAGM, Manoe R, Muller A, Visser CA, Meijer CJLM, Musters RJP, Hack CE, Aarden LA, and Niessen HWM: Clusterin: a protective mediator for ischemic cardiomyocytes? *Am J Physiol Heart Circ Physiol* 2005; 289:H2193-H2202


- Kim BM, Kim SY, Lee S, Shin YJ, Min BH, Bendayan M, Park IS: Clusterin induces differentiation of pancreatic duct cells into insulin-secreting cells. *Diabetologia* 2006; 49:311-320
- Kruger S, Mahnken A, Kausch I, Feller AC: Value of Clusterin immunoreactivity as a predictive factor in muscle-invasive urothelial bladder carcinoma. *Urology* 2006; 67:105-109
- Rodriguez-Pineiro AM, De la Cadena MP, Lopez-Saco A, and Rodriguez-Berrocal FJ: Differential Expression of serum clusterin isoforms in colorectal cancer. *Mol. And Cel. Proteomics* 2006; 5:1647-1657
- Strochi P, Smith MA, Perry G, Tamagno E, Danni O, Pession A, Gaiba A, Dozza B: Clusterin up-regulation following sub-lethal oxidative stress and lipid peroxidation in human neuroblastoma cells. *Neurobiol. of Aging* 2006; 27:1588-1594
- Ishii A, Sakai Y, and Nakamura A: Molecular pathological evaluation of clusterin in a rat model of unilateral ureteral obstruction as a possible biomarker of nephrotoxicity. *Toxicologic Pathology* 2007; 35:376-382
- Stoop MP, Dekker LJ, Titulaer MK, Burgers PC, Sillevius Smitt PAE, Luiders TM, and Hintzen RQ: Multiple sclerosis-related identified in cerebrospinal fluid by advanced mass spectrometry. *Proteomics* 2008; 8:0000-0000

➤➤ For more references on this product see our WebPages at www.biovendor.com

18. EXPLANATION OF SYMBOLS


	Catalogue number
	Content
	Lot number
	Attention, see instructions for use
	Potential biological hazard
	Expiry date
	Storage conditions
	Name and registered office of the manufacturer

Assay Procedure Summary





12								
11								
10								
9								
8								
7								
6								
5								
4								
3								
2								
1								
	A	B	C	D	E	F	G	H

NOTES



NOTES





BioVendor – Laboratorní medicína a.s.

Karasek 1767/1, 621 00 Brno, Czech Republic

Phone: +420-549-124-185, Fax: +420-549-211-460

E-mail: info@biovendor.com, sales@biovendor.com

Web: www.biovendor.com

There are BioVendor branches and distributors near you.

To find the office closest to you, visit www.biovendor.com/contact

