

HP IgA

**Enzyme ImmunoAssay (ELISA) for
the quantitative/qualitative
determination of IgA antibodies to
Helicobacter pylori
in human serum and plasma**

- for "in vitro" diagnostic use only -



DIA.PRO

**Diagnostic Bioprobes Srl
Via G. Carducci n° 27
20099 Sesto San Giovanni
(Milano) - Italy**

Phone +39 02 27007161

Fax +39 02 26007726

e-mail: info@diapro.it

HP IgA

A. INTENDED USE

Enzyme ImmunoAssay (ELISA) for the quantitative/qualitative determination of IgA antibodies to Helicobacter pylori in human plasma and sera. The product is intended for the follow-up of patients showing gastrointestinal pathologies referable to H.pylori infection.

For "in vitro" diagnostic use only.

B. INTRODUCTION

Helicobacter pylori (HP) is a Gram negative bacterium, firstly isolated in gastric mucosa by Marshall and Warren in 1983.

Hp has been recognized to be the agent responsible of most of cases of gastric mucosal damage and to play a role in the evolution of gastric diseases to carcinoma.

Hp causes an immunological response during infection and specific antibodies of the different classes of IgG, IgA and IgM are produced by the patient.

ELISA are currently used to screen patients affected by gastritis or peptic ulcers for acute active infection due to some Helicobacter pylori virulent strains.

In particular the presence of IgA and IgM antibodies is reported to be correlated to the acute phase of illness, while IgG antibodies become present at different titers shortly after primary infections and last in blood for many years.

Quantitative ELISA are also used in the follow-up of patients undergoing antibiotic therapy, useful in monitoring IgG titer variations during and after the pharmaceutical treatment

C. PRINCIPLE OF THE TEST

Microplates are coated with H.pylori immunodominant antigens derived from tissue culture of a virulent strain.

In the 1st incubation, the solid phase is treated with diluted samples and anti-HP IgA are captured, if present, by the antigens.

After washing out all the other components of the sample, in the 2nd incubation bound anti-HP IgA are detected by the addition of anti hlgA antibody, labeled with peroxidase (HRP).

The enzyme captured on the solid phase, acting on the substrate/chromogen mixture, generates an optical signal that is proportional to the amount of anti-HP IgA antibodies present in the sample.

IgA in the sample may therefore be quantitated by means of a standard curve calibrated in arbitrary units per milliliter (arbU/ml) as no international standard is available.

Neutralization of IgG anti-HP, carried out directly in the well, is performed in the assay in order to block interferences due to this class in the determination of IgA.

D. COMPONENTS

Each kit contains sufficient reagents to perform 96 tests.

1. Microplate: MICROPLATE

12 strips x 8 microwells coated with HP specific immunodominant antigens derived from tissue culture of a virulent strain. Plates are sealed into a bag with desiccant.

Allow the microplate to reach room temperature before opening; reseal unused strips in the bag with desiccant and store at 4°C.

2. Calibration Curve: CAL N°...

6x2.0 ml/vial. Ready to use and color coded standard curve ranging: CAL1 = 0 arbU/ml // CAL2 = 5 arbU/ml // CAL3 = 10 arbU/ml // CAL4 = 20 arbU/ml // CAL 5 = 50 arbU/ml // CAL6 = 100 arbU/ml. Standards are calibrated against an internal Gold Standard or IGS as no international one is defined.

Contains human serum proteins, 2% casein, 10 mM Na-citrate buffer pH 6.0 +/-0.1, 0.1% Tween 20, 0.09% Na-azide and 0.1% Kathon GC as preservatives. Standards are blue color coded.

3. Control Serum: CONTROL ...ml

1 vial. Lyophilized. It contains fetal bovine serum proteins, human IgA antibodies to HP at about 20 arbU/ml±20%, 0.3 mg/ml gentamicine sulphate and 0.1% Kathon GC as preservatives.

4. Wash buffer concentrate: WASHBUF 20X

1x60ml/bottle20x concentrated solution.

Once diluted, the wash solution contains 10 mM phosphate buffer pH 7.0+/-0.2, 0.05% Tween 20 and 0.05% Kathon GC.

5. Enzyme conjugate: CONJ

1x16ml/vial. Ready to use and red colour coded. It contains Horseradish peroxidase conjugated polyclonal antibodies to human IgA, 5% BSA, 10 mM Tris buffer pH 6.8+/-0.1, 0.1% Kathon GC and 0.02% gentamicine sulphate as preservatives.

6. Chromogen/Substrate: SUBS TMB

1x16ml/vial. It contains 50 mM citrate-phosphate buffer pH 3.5-3.8, 4% dimethylsulphoxide, 0.03% tetra-methyl-benzidine (or TMB) and 0.02% hydrogen peroxide (or H₂O₂).

Note: To be stored protected from light as sensitive to strong illumination.

7. Sulphuric Acid: H₂SO₄ 0.3 M

1x15ml/vialll contains 0.3 M H₂SO₄ solution.

Attention !: Irritant (Xi R36/38; S2/26/30)

8. Specimen Diluent: DILSPE

2x60ml/vial. It contains 2% casein, 10 mM Na-citrate buffer pH 6.0 +/-0.1, 0.1% Tween 20, 0.09% Na-azide and 0.1% Kathon GC as preservatives. To be used to dilute the sample.

9. Neutralizing Reagent: SOLN NEUT

1x8ml/vial. Ready-to-use Reagent. It contains goat anti hlgG, 2% casein, 10 mM Na-citrate buffer pH 6.0 +/-0.1, 0.1% Tween 20, 0.09% Na-azide and 0.1% Kathon GC as preservatives.

10. Plate sealing foils n°2

11. Package insert n°1

E. MATERIALS REQUIRED BUT NOT PROVIDED

1. Calibrated Micropipettes (1000, 100 and 10ul) and disposable plastic tips.
2. EIA grade water (bidistilled or deionised, charcoal treated to remove oxidizing chemicals used as disinfectants).
3. Timer with 60 minute range or higher.
4. Absorbent paper tissues.
5. Calibrated ELISA microplate thermostatic incubator (dry or wet) set at +37°C (+/-0.5°C tolerance).
6. Calibrated ELISA microwell reader with 450nm (reading) and possibly with 620-630nm (blinking) filters.
7. Calibrated ELISA microplate washer.
8. Vortex or similar mixing tools.

F. WARNINGS AND PRECAUTIONS

1. The kit has to be used by skilled and properly trained technical personnel only, under the supervision of a medical doctor responsible of the laboratory.

2. All the personnel involved in performing the assay have to wear protective laboratory clothes, talc-free gloves and glasses. The use of any sharp (needles) or cutting (blades) devices should be avoided. All the personnel involved should be trained

in biosafety procedures, as recommended by the Center for Disease Control, Atlanta, U.S. and reported in the National Institute of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.

3. All the personnel involved in sample handling should be vaccinated for HBV and HAV, for which vaccines are available, safe and effective.
4. The laboratory environment should be controlled so as to avoid contaminants such as dust or air-borne microbial agents, when opening kit vials and microplates and when performing the test. Protect the Chromogen (TMB) from strong light and avoid vibration of the bench surface where the test is undertaken.
5. Upon receipt, store the kit at 2..8°C into a temperature controlled refrigerator or cold room.
6. Do not interchange components between different lots of the kits. It is recommended that components between two kits of the same lot should not be interchanged.
7. Check that the reagents are clear and do not contain visible heavy particles or aggregates. If not, advise the laboratory supervisor to initiate the necessary procedures for kit replacement.
8. Avoid cross-contamination between serum/plasma samples by using disposable tips and changing them after each sample.
9. Avoid cross-contamination between kit reagents by using disposable tips and changing them between the use of each one.
10. Do not use the kit after the expiration date stated on the external container and internal (vials) labels. A study conducted on an opened kit did not pointed out any relevant loss of activity up to six 6 uses of the device and up to 6 months.
11. Treat all specimens as potentially infective. All human serum specimens should be handled at Biosafety Level 2, as recommended by the Center for Disease Control, Atlanta, U.S. in compliance with what reported in the Institutes of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.
12. The use of disposable plastic-ware is recommended in the preparation of the liquid components or in transferring components into automated workstations, in order to avoid cross contamination.
13. Waste produced during the use of the kit has to be discarded in compliance with national directives and laws concerning laboratory waste of chemical and biological substances. In particular, liquid waste generated from the washing procedure, from residuals of controls and from samples has to be treated as potentially infective material and inactivated before waste. Suggested procedures of inactivation are treatment with a 10% final concentration of household bleach for 16-18 hrs or heat inactivation by autoclave at 121°C for 20 min..
14. Accidental spills from samples and operations have to be adsorbed with paper tissues soaked with household bleach and then with water. Tissues should then be discarded in proper containers designated for laboratory/hospital waste.
15. The Sulphuric Acid is an irritant. In case of spills, wash the surface with plenty of water
16. Other waste materials generated from the use of the kit (example: tips used for samples and controls, used microplates) should be handled as potentially infective and disposed according to national directives and laws concerning laboratory wastes.

G. SPECIMEN: PREPARATION AND WARNINGS

1. Blood is drawn aseptically by venepuncture and plasma or serum is prepared using standard techniques of preparation of samples for clinical laboratory analysis. No influence has been observed in the preparation of the sample with citrate, EDTA and heparin.
2. Samples have to be clearly identified with codes or names in order to avoid misinterpretation of results. Bar code labeling and electronic reading is strongly recommended.

3. Haemolysed ("red") and visibly hyperlipemic ("milky") samples have to be discarded as they could generate false results. Samples containing residues of fibrin or heavy particles or microbial filaments and bodies should be discarded as they could give rise to false results.

4. Sera and plasma can be stored at +2°.8°C for up to five days after collection. For longer storage periods, samples can be stored frozen at -20°C for several months. Any frozen samples should not be freeze/thawed more than once as this may generate particles that could affect the test result.

5. If particles are present, centrifuge at 2.000 rpm for 20 min or filter using 0.2-0.8µ filters to clean up the sample for testing.

H. PREPARATION OF COMPONENTS AND WARNINGS

Microplate:

Allow the microplate to reach room temperature (about 1 hr) before opening the container. Check that the desiccant is not turned to dark green, indicating a defect of storage.

In this case call Dia.Pro's customer service.

Unused strips have to be placed back into the aluminium pouch, in presence of desiccant supplied, firmly zipped and stored at +2°.8°C. When opened the first time, residual strips are stable till the indicator of humidity inside the desiccant bag turns from yellow to green.

Calibration Curve

Ready to use component. Mix carefully on vortex before use.

Control Serum

Add the volume of ELISA grade water, reported on the label, to the lyophilised powder; let fully dissolve and then gently mix on vortex.

Note: *The control after dissolution is not stable. Store frozen in aliquots at -20°C.*

Wash buffer concentrate:

The whole content of the concentrated solution has to be diluted 20x with bidistilled water and mixed gently end-over-end before use. During preparation avoid foaming as the presence of bubbles could impact on the efficiency of the washing cycles.

Note: *Once diluted, the wash solution is stable for 1 week at +2..8°C.*

Enzyme conjugate:

Ready to use. Mix well on vortex before use.

Be careful not to contaminate the liquid with oxidizing chemicals, air-driven dust or microbes.

If this component has to be transferred use only plastic, possibly sterile disposable containers.

Chromogen/Substrate:

Ready to use. Mix well on vortex before use.

Be careful not to contaminate the liquid with oxidizing chemicals, air-driven dust or microbes.

Do not expose to strong illumination, oxidizing agents and metallic surfaces.

If this component has to be transferred use only plastic, possible sterile disposable container

Sample Diluent

Ready to use component. Mix carefully on vortex before use.

Neutralizing Reagent

Ready to use component. Mix carefully on vortex before use.

Sulphuric Acid:

Ready to use. Mix well on vortex before use.

Legenda: R 36/38 = Irritating to eyes and skin.

S 2/26/30 = In case of contact with eyes, rinse immediately with plenty of water and seek medical advice.

I. INSTRUMENTS AND TOOLS USED IN COMBINATION WITH THE KIT

1. Micropipettes have to be calibrated to deliver the correct volume required by the assay and must be submitted to regular decontamination (household alcohol, 10% solution of bleach, hospital grade disinfectants) of those parts that could accidentally come in contact with the sample. They should also be regularly maintained in order to show a precision of 1% and a trueness of +/-2%. Decontamination of spills or residues of kit components should also be carried out regularly.
2. The ELISA incubator has to be set at +37°C (tolerance of +/-0.5°C) and regularly checked to ensure the correct temperature is maintained. Both dry incubators and water baths are suitable for the incubations, provided that the instrument is validated for the incubation of ELISA tests.
3. The ELISA washer is extremely important to the overall performances of the assay. The washer must be carefully validated and correctly optimised using the kit controls and reference panels, before using the kit for routine laboratory tests. Usually 4-5 washing cycles (aspiration + dispensation of 350ul/well of washing solution = 1 cycle) are sufficient to ensure that the assay performs as expected. A soaking time of 20-30 seconds between cycles is suggested. In order to set correctly their number, it is recommended to run an assay with the kit controls and well characterized negative and positive reference samples, and check to match the values reported below in the section "Internal Quality Control". Regular calibration of the volumes delivered by, and maintenance (decontamination and cleaning of needles) of the washer has to be carried out according to the instructions of the manufacturer.
4. Incubation times have a tolerance of ±5%.
5. The ELISA microplate reader has to be equipped with a reading filter of 450nm and ideally with a second filter (620-630nm) for blanking purposes. Its standard performances should be (a) bandwidth ≤ 10 nm; (b) absorbance range from 0 to ≥ 2.0; (c) linearity to ≥ 2.0; repeatability ≥ 1%. Blanking is carried out on the well identified in the section "Assay Procedure". The optical system of the reader has to be calibrated regularly to ensure that the correct optical density is measured. It should be regularly maintained according to the manufacturer's instructions.
6. When using an ELISA automated work station, all critical steps (dispensation, incubation, washing, reading, data handling) have to be carefully set, calibrated, controlled and regularly serviced in order to match the values reported in the section "Internal Quality Control". The assay protocol has to be installed in the operating system of the unit and validated as for the washer and the reader. In addition, the liquid handling part of the station (dispensation and washing) has to be validated and correctly set. Particular attention must be paid to avoid carry over by the needles used for dispensing and for washing. This must be studied and controlled to minimize the possibility of contamination of adjacent wells. The use of ELISA automated work stations is recommended when the number of samples to be tested exceed 20-30 units per run.
7. Dia.Pro's customer service offers support to the user in the setting and checking of instruments used in combination with the kit, in order to assure compliance with the requirements described. Support is also provided for the installation of new instruments to be used with the kit.

L. PRE ASSAY CONTROLS AND OPERATIONS

1. Check the expiration date of the kit printed on the external label (primary container). Do not use if expired.
2. Check that the liquid components are not contaminated by visible particles or aggregates.

3. Check that the Chromogen (TMB) is colourless or pale blue by aspirating a small volume of it with a sterile plastic pipette.
4. Check that no breakage occurred in transportation and no spillage of liquid is present inside the box (primary container). Check that the aluminium pouch, containing the microplate, is not punctured or damaged.
5. Dissolve the content of the Control Serum as reported.
6. Dilute all the content of the 20x concentrated Wash Solution as described above.
7. Allow all the other components to reach room temperature (about 1 hr) and then mix gently on vortex all liquid reagents.
8. Set the ELISA incubator at +37°C and prepare the ELISA washer by priming with the diluted washing solution, according to the manufacturer's instructions. Set the right number of washing cycles as found in the validation of the instrument for its use with the kit.
9. Check that the ELISA reader is turned on or ensure it will be turned on at least 20 minutes before reading.
10. If using an automated work station, turn on, check settings and be sure to use the right assay protocol.
11. Check that the micropipettes are set to the required volume.
12. Check that all the other equipment is available and ready to use.
13. In case of problems, do not proceed further with the test and advise the supervisor.

M. ASSAY PROCEDURE

The assay has to be carried out according to what reported below, taking care to maintain the same incubation time for all the samples in testing.

The kit may be used for quantitative and qualitative determinations as well.

M1. QUANTITATIVE DETERMINATION:

1. Dilute samples 1:101 into a properly defined dilution tube (example: 1000 µl Sample Diluent + 10 µl sample). Do not dilute the Calibration Set as calibrators are ready to use. Mix carefully all the liquid components on vortex and then proceed as described below.
2. Place the required number of Microwells in the microwell holder. Leave the A1 and B1 empty for the operation of blanking.
3. Dispense 50 µl of the Neutralizing Reagent (SOLN NTR) in all the wells of the samples. Do not add it in the wells used for the Calibrators and the Control Serum!
4. Then dispense 100 µl of Calibrators and 100 µl Control Serum in duplicate. Then dispense 100 µl of diluted samples in each properly identified well.
5. Incubate the microplate for **60 min at +37°C**.

Important note: Strips have to be sealed with the adhesive sealing foil, supplied, only when the test is carried out manually. Do not cover strips when using ELISA automatic instruments.

6. Wash the microplate with an automatic washer as reported previously (section I.3).
7. Pipette 100 µl Enzyme Conjugate into each well, except A1+B1 blanking wells, and cover with the sealer. Check that this red coloured component has been dispensed in all the wells, except A1 and B1.

Important note: Be careful not to touch the plastic inner surface of the well with the tip filled with the Enzyme Conjugate. Contamination might occur.

8. Incubate the microplate for **60 min at +37°C**.

- Wash microwells as in step 5.
- Pipette 100 µl Chromogen/Substrate mixture into each well, the blank wells A1 and B1 included. Then incubate the microplate at **room temperature (18-24°C) for 20 minutes**.

Important note: Do not expose to strong direct illumination. High background might be generated.

- Pipette 100 µl Sulphuric Acid to stop the enzymatic reaction into all the wells using the same pipetting sequence as in step 9. Addition of acid will turn the positive calibrators, the control serum and the positive samples from blue to yellow.
- Measure the colour intensity of the solution in each well, as described in section I.5, at 450nm filter (reading) and possibly at 620-630nm (background subtraction), blanking the instrument on A1 or B1 or both.

M2. QUALITATIVE DETERMINATION

If only a qualitative determination is required, proceed as described below:

- Dilute samples 1:101 into a properly defined dilution tube (example: 1000 µl Sample Diluent + 10 µl sample). Do not dilute the Calibration Set as calibrators are ready to use. Mix carefully all the liquid components on vortex and then proceed as described below.
- Place the required number of Microwells in the microwell holder. Leave A1 well empty for the operation of blanking.
- Dispense 50 µl of the Neutralizing Reagent (SOLN NTR) in all the wells of the samples. Do not add it in the wells used for the Calibrators !
- Dispense 100 µl of Calibrator 0 arbU/ml and Calibrator 5 arbU/ml in duplicate and Calibrator 100 arbU/ml in single. Then dispense 100 µl of diluted samples in each properly identified well.
- Incubate the microplate for **60 min at +37°C**.

Important note: Strips have to be sealed with the adhesive sealing foil, supplied, only when the test is carried out manually. Do not cover strips when using ELISA automatic instruments.

- Wash the microplate with an automatic as reported previously (section I.3).
- Pipette 100 µl Enzyme Conjugate into each well, except the A1 well, and cover with the sealer. Check that this red coloured component has been dispensed in all the wells, except A1.

Important note: Be careful not to touch the plastic inner surface of the well with the tip filled with the Enzyme Conjugate. Contamination might occur.

- Incubate the microplate for **60 min at +37°C**.
- Wash microwells as in step 6.
- Pipette 100 µl Chromogen/Substrate mixture into each well, the blank well included. Then incubate the microplate at **room temperature (18-24°C) for 20 minutes**.

Important note: Do not expose to strong direct illumination. High background might be generated.

- Pipette 100 µl Sulphuric Acid into all the wells using the same pipetting sequence as in step 9. Addition of acid will turn the positive calibrators, the control serum and the positive samples from blue to yellow.
- Measure the colour intensity of the solution in each well, as described in section I.5, at 450nm filter (reading) and possibly at 620-630nm (background subtraction), blanking the instrument on A1.

General Important notes:

- If the second filter is not available ensure that no finger prints are present on the bottom of the microwell before reading at 450nm. Finger prints could generate false positive results on reading.
- Reading has to be carried out just after the addition of the Stop Solution and anyway not any longer than 20 minutes after its addition. Some self oxidation of the chromogen can occur leading to high background.

N. ASSAY SCHEME

Method	Operations
Neutralizing Reagent	50 µl
Calibrators & Control	100 µl
Samples diluted 1:101	100 µl
1st incubation	60 min
Temperature	+37°C
Wash step	4-5 cycles
Enzyme conjugate	100 µl
2nd incubation	60 min
Temperature	+37°C
Wash step	4-5 cycles
TMB/H ₂ O ₂	100 µl
3rd incubation	20 min
Temperature	r.t.
Sulphuric Acid	100 µl
Reading OD	450nm

An example of dispensation scheme for Quantitative Analysis is reported below:

Microplate

	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	CAL4	S 1									
B	BLK	CAL4	S 2									
C	CAL1	CAL5	S 3									
D	CAL1	CAL5	S 4									
E	CAL2	CAL6	S 5									
F	CAL2	CAL6	S 6									
G	CAL3	CS	S 7									
H	CAL3	CS	S 8									

Legenda: BLK = Blank CAL = Calibrator
CS = Control Serum S = Sample

An example of dispensation scheme in qualitative assays is reported below:

Microplate

	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	S 3	S 11									
B	CAL1	S 4	S 12									
C	CAL1	S 5	S 13									
D	CAL2	S 6	S 14									
E	CAL2	S 7	S 15									
F	CAL6	S 8	S 16									
G	S 1	S 9	S 17									
H	S 2	S 10	S 18									

Legenda: BLK = Blank CAL = Calibrators
S = Sample

O. INTERNAL QUALITY CONTROL

A validation check is carried out on the controls any time the kit is used in order to verify whether the performances of the assay are as qualified.

Control that the following data are matched:

Check	Requirements
Blank well	< 0.100 OD450nm value
CAL 1 0 arbU/ml	< 0.150 mean OD450nm value after blanking coefficient of variation < 30%
CAL 2 5 arbU/ml	OD450nm > OD450nm CAL1 + 0.100
CAL 6 100 arbU/ml	OD450nm > 1.000
Control Serum	20 arbU/ml +/-20%

If the results of the test match the requirements stated above, proceed to the next section.

If they do not, do not proceed any further and operate as follows:

Problem	Check
Blank well > 0.100 OD450nm	1. that the Chromogen/Sustrate solution has not got contaminated during the assay
CAL 1 0 arbU/ml > 0.150 OD450nm after blanking coefficient of variation > 30%	1. that the washing procedure and the washer settings are as validated in the pre qualification study; 2. that the proper washing solution has been used and the washer has been primed with it before use; 3. that no mistake has been done in the assay procedure (dispensation of a positive calibrator instead of the negative one); 4. that no contamination of the negative calibrator or of their wells has occurred due spills of positive samples or the enzyme conjugate; 5. that micropipettes haven't got contaminated with positive samples or with the enzyme conjugate 6. that the washer needles are not blocked or partially obstructed.
CAL 2 5 arbU/ml OD450nm < OD450nm CAL1 + 0.100	1. that the procedure has been correctly executed; 2. that no mistake has been done in its distribution (ex.: dispensation of a wrong calibrator instead); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the calibrator has occurred.
CAL 6 100 arbU/ml < 1.000 OD450nm	1. that the procedure has been correctly executed; 2. that no mistake has been done in its distribution (dispensation of a wrong calibrator instead); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the positive control has occurred.
Control Serum Different from expected value	1. that the procedure has been correctly executed; 2. that no mistake has been done in its distribution (dispensation of a wrong calibrator instead); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the positive control has occurred.

Should one of these problems have happened, after checking, report to the supervisor for further actions.


P. RESULTS

P.1 Quantitative method

If the test turns out to be valid, use for the quantitative method an approved curve fitting program to draw the calibration curve from the values obtained by reading at 450nm (4-parameters interpolation is suggested).

Then on the calibration curve calculate the concentration of anti H.pylori IgA antibody in samples.

An example of Calibration curve is reported below.



Important Note:

Do not use the calibration curve above to make calculations.

P.2 Qualitative method

In the qualitative method, calculate the mean OD450nm values for the Calibrators 0 and 5 arbU/ml and then check that the assay is valid.

Example of calculation:

Note: The following data must not be used instead of real figures obtained by the user.

Calibrator 0 arbU/ml: 0.020 – 0.024 OD450nm
Mean Value: 0.022 OD450nm
Lower than 0.150 – Accepted

Calibrator 5 arbU/ml: 0.250 – 0.270 OD450nm
Mean Value: 0.260 OD450nm
Higher than Cal 0 + 0.100 – Accepted

Calibrator 100 arbU/ml: 2.045 OD450nm
Higher than 1.000 – Accepted

The OD450nm of the Calibrator 5 arbU/ml is considered the cut-off (or Co) of the system.

The ratio between the OD450nm value of the sample and the OD450nm of the Calibrator 5 arbU/ml (or S/Co) can provide a semi-quantitative estimation of the content of specific IgG in the sample.

Q. INTERPRETATION OF RESULTS

Samples with a concentration lower than 5 arbU/ml are considered negative for anti H.pylori IgA antibody. Samples with a concentration higher than 5 arbU/ml are considered positive for anti H.pylori IgA antibody.

Important notes:

1. *H.pylori IgA results alone are not enough to provide a clear diagnosis of Helicobacter pylori infection. Other tests for Helicobacter pylori (supplied by Dia.Pro Diagnostic BioProbes s.r.l. at code n° HPAG.CE, HPG.CE and HPM.CE), should be carried out.*
2. *Interpretation of results should be done under the supervision of the laboratory supervisor to reduce the risk of judgment errors and misinterpretations.*
3. *When test results are transmitted from the laboratory to another facility, attention must be paid to avoid erroneous data transfer.*
4. *Diagnosis has to be done and released to the patient by a suitably qualified medical doctor.*

R. PERFORMANCE CHARACTERISTICS

Evaluation of Performances has been conducted on panels of positive and negative samples in an external clinical laboratory with reference to a FDA approved reference kit.

1. Limit of detection

No international standard for HP IgA Antibody detection has been defined so far by the European Community. In its absence, an Internal Gold Standard (or IGS), derived from a patient with an history of past HP infection, has been defined in order to provide the device with a constant and excellent sensitivity.

2. Diagnostic Sensitivity and Specificity:

The diagnostic performances were evaluated in a performance evaluation study conducted in an external center, with excellent experience in the diagnosis of infectious diseases.

The **diagnostic sensitivity** was studied in an external study on more than 50 samples, pre-tested positive with the reference kit of European origin in use at the laboratory. Positive samples were collected from patients with a clinical history of H.pylori infection.

The **diagnostic specificity** was determined in an external study on panels of more than 100 negative samples from normal individuals and blood donors, classified negative with the reference kit, including potentially interfering specimens.

Both plasma, derived with different standard techniques of preparation (citrate, EDTA and heparin), and sera have been used to determine the specificity. No false reactivity due to the method of specimen preparation has been observed.

In addition 35 cross-reacting specimens were studied internally to verify absence of interference on the assay results.

No interference was observed (100% specificity).

Frozen specimens have also been tested to check whether samples freezing interferes with the performance of the test. No interference was observed on clean and particle free samples.

The external Performance Evaluation provided the following values :

Sensitivity	> 98 %
Specificity	> 98 %

3. Reproducibility:

A study conducted on three samples of different HP IgA reactivity, examined in 16 replicates in three separate runs has

shown CV% values ranging 7-18% depending on the OD450nm readings.

The variability shown in the tables did not result in sample misclassification.

S. LIMITATIONS

False positivity has been assessed as less than 2% of the normal population.

Frozen samples containing fibrin particles or aggregates may generate false positive results.

REFERENCES

1. Lazzaroni M. et al.. Medicina (1989), 9, 9-18.
2. Vaira D. et al.. Federazione Medica XLI (1988), 7, 549-555.
3. Oderda G. Et al.. The Lancet (1989), vol.6, 7, 358-360.
4. Loffeld H. et al.. The Lancet (1989) vol.6, 10, 554-556
5. Vaira D. et al.. British Medical Journal (1988), vol.9, 43, 374-375.
6. Oderda G. et a.. Gut (1989), vol. 30, 7, 912-916.
7. Vaira D. et al.. Ital.J.Gastroenterol. (1988), 20, 299-304.
8. Vaira D. et al.. Current Opinion in Gastroenterology (1989), 5, 817-823.

<p>Produced by Dia.Pro. Diagnostic Bioprobes Srl. Via G. Carducci n°27 – Sesto San Giovanni (MI) - Ita ly</p>

