



Instruction Manual

Mumps IgG ELISA

Enzyme immunoassay based on microtiter plate
for the detection and quantitative determination
of human IgG antibodies against **Mumps Virus**
in serum and plasma



Cat. No.: ILE-MUM01
Storage: 2-8°C
For in-vitro diagnostic use only

September 2010

Contents	Page
1. Intended Use	3
2. General Information	3
3. Principle of the Test	3
4. Limitations, Precautions and General Comments	4
5. Reagents Provided	4
6. Materials Required but not Provided	5
7. Specimen Collection and Handling	6
8. Assay Procedure	6
9. Evaluation	7
10. Assay Characteristics	8
11. References	8

Symbole und Übersetzungen / Symbols and Translations

Symbol	English	French	German	Italian	Spanish	Greek
CAL	Calibrator	Etalon	Kalibrator	Calibratore	Calibrador	Πρότυπο Διάλυμα
CONJ	Conjugate	Conjugué	Konjugat	Coniugato	Conjugado	Διάλυμα Συμπλόκου
CONC	Concentrate (<n>-fold)	Concentré (<n> fois)	Konzentrat (<n>-fach)	Concentrato (<n>-volte)	Concentrado (<n>-veces)	Συμπύκνωση (<n> φορές)
SAMP DIL	Sample Diluent	Diluant échantillon	Proben- verdünner	Diluyente del campione	Diluyente de muestra	Διάλυμα Αραίωσης Δειγμάτων
STOP	Stop Solution	Solution d'arrêt	Stopp-Lösung	Soluzione d'arresto	Solución de parada	Διάλυμα Αναστολής
SUBS	Substrate	Substrat	Substrat	Substrato	Sustrato	Διάλυμα Υποστρώ- ματος
MT PLATE	Microtiter plate	Microplaque	Mikrotiterplatte	Piastre	Placa microtiter	Μικρόπλακα
WASH BUF	Wash buffer	Tampon de lavage	Waschpuffer	Soluzione di lavaggio	Tampón de lavado	Πλυστικό Διάλυμα

1. Intended Use

The IMMUNOLAB Mumps IgG Antibody ELISA Test Kit has been designed for the detection and the quantitative determination of specific IgG antibodies against Mumps virus in serum and plasma. Further applications in other body fluids are possible and can be requested from the Technical Service of IMMUNOLAB.

This assay is intended for in-vitro diagnostic use only.

Laboratory results can never be the only base of a medical report. The patient history and further tests have additionally to be taken into account.

2. General Information

Mumps (Parotitis) is a common contagious disease with relatively moderate symptoms during childhood, but increasing complications, when adults are infected. The causative agent of mumps is a virus of the Paramyxoviridae family. The virus normally infects children at the age of 4 to 10. The infection is mainly transmitted by the airborne route, but is also spread by various objects contaminated by the patient saliva. The disease shows a seasonal prevalence with the greatest incidence in winter and spring. Both a mumps infection or vaccination lead to a persistent immunity.

The typical symptom associated with a mumps infection is a „parotitis“ (swelling of the parotid glands). Additionally pathological involvement of the CNS and various glandular organs (pancreas, thymus, thyroids) is a typical feature of mumps. A mumps induced meningitis is one of the most frequent manifestations of the disease which can also appear without parotitis. The published data about the presence of meningitis vary between 1.4% and 66% of the clinically ill patients. In most cases there follows reconvalescence without complications.

The recognition of Mumps in the laboratory is mostly done by detection of the infectious agent itself or by the determination of virus-specific antibodies. The serodiagnosis plays a major part. Besides the classical methods like complement fixation, hemagglutination and neutralisation tests have been introduced a series of modern assays like immunofluorescence, radio immunoassay and enzyme immunoassay. The immunity of an individual is monitored by an IgG ELISA. Increasing IgG titers are helpful in the determination of the causative agent because cross reactions with Parainfluenza-Virus type 2 may lead to a wrong interpretation of the patient reports. In the early stage of the disease, significant IgM titers may be monitored.

3. Principle of the Test

The IMMUNOLAB Mumps IgG antibody test kit is based on the principle of the enzyme immunoassay (EIA). Mumps antigen is bound on the surface of the microtiter strips. Diluted patient serum or ready-to-use standards are pipetted into the wells of the microtiter plate. A binding between the IgG antibodies of the serum and the immobilized Mumps antigen takes place. After a one hour incubation at room temperature, the plate is rinsed with diluted wash solution, in order to remove unbound material. Then ready-to-use anti-human-IgG peroxidase conjugate is added and incubated for 30 minutes. After a further washing step, the substrate (TMB) solution is pipetted and incubated for 20 minutes, inducing the development of a blue dye in the wells. The color development is terminated by the addition of a stop solution, which changes the color from blue to yellow. The resulting dye is measured spectrophotometrically at the wavelength of 450 nm. The concentration of the IgG antibodies is directly proportional to the intensity of the color.

4. Limitations, Precautions and General Comments

- Only for in-vitro use! Do not ingest or swallow! The usual laboratory safety precautions as well as the prohibition of eating, drinking and smoking in the lab have to be followed.
- All sera and plasma or buffers based upon, have been tested respective to HBsAg, HIV and HCV with recognized methods and were found negative. Nevertheless precautions like the use of latex gloves have to be taken.
- Serum and reagent spills have to be wiped off with a disinfecting solution (e.g. sodium hypochlorite, 5%) and have to be disposed of properly.
- All reagents have to be brought to room temperature (18 to 25 °C) before performing the test.
- Before pipetting all reagents should be mixed thoroughly by gentle tilting or swinging. Vigorous shaking with formation of foam should be avoided.
- It is important to pipet with constant intervals, so that all the wells of the microtiter plate have the same conditions.
- When removing reagents out of the bottles, care has to be taken that the stoppers are not contaminated. Further a possible mix-up has to be avoided. The content of the bottles is usually sensitive to oxidation, so that they should be opened only for a short time.
- In order to avoid a carry-over or a cross-contamination, separate disposable pipet tips have to be used.
- No reagents from different kit lots have to be used, they should not be mixed among one another.
- All reagents have to be used within the expiry period.
- In accordance with a Good Laboratory Practice (GLP) or following ISO9001 all laboratory devices employed should be regularly checked regarding the accuracy and precision. This refers amongst others to microliter pipets and washing or reading (ELISA-Reader) instrumentation.
- The contact of certain reagents, above all the stopping solution and the substrate with skin, eye and mucosa has to be avoided, because possible irritations and acid burns could arise, and there exists a danger of intoxication.

5. Reagents Provided

Store kit components at 2-8°C and do not use after the expiry date on the box outer label. Before use, all components should be allowed to warm up to ambient temperature (18-25°C). After use, the plate should be resealed, the bottle caps replaced and tightened and the kit stored at 2-8°C. The opened kit should be used within three months.

Components	Volume / Qty.
Mumps antigen coated microtiter strips	12
Calibrator A (Negative Control)	2 mL
Calibrator B (Cut-Off Standard)	2 mL
Calibrator C (Weak Positive Control)	2 mL
Calibrator D (Positive Control)	2 mL
Enzyme Conjugate	15 mL
Substrate	15 mL
Stop Solution	15 mL
Sample Diluent	60 mL
Washing Buffer (10×)	60 mL
Plastic foils	2
Plastic bag	1

5.1. Microtiter Strips

12 strips with 8 breakable wells each, coated with a Mumps antigen (Strain Enders, ATCC VR-106). Ready-to-use.

5.2. Calibrator A (Negative Control)

2 mL, protein solution diluted with PBS, contains no IgG antibodies against Mumps. Addition of 0.01 % methylisothiazolone and 0.01 % bromonitrodioxane. Ready-to-use.

5.3. Calibrator B (Cut-Off Standard)

2 mL human serum diluted with PBS, contains a low concentration of IgG antibodies against Mumps. Addition of 0.01 % methylisothiazolone and 0.01 % bromonitrodioxane. Ready-to-use.

5.4. Calibrator C (Weak Positive Control)

2 mL, human serum diluted with PBS, contains a medium concentration of IgG antibodies against Mumps. Addition of 0.01 % methylisothiazolone and 0.01 % bromonitrodioxane. Ready-to-use.

5.5. Calibrator D (Positive Control)

2 mL, human serum diluted with PBS, contains a high concentration of IgG antibodies against Mumps. Addition of 0.01 % methylisothiazolone and 0.01 % bromonitrodioxane. Ready-to-use.

5.6. Enzyme Conjugate

15 mL, anti-human-IgG-HRP (rabbit), in protein-containing buffer solution. Addition of 0.01 % methylisothiazolone and 0.01 % bromonitrodioxane and 5 mg/L ProclinTM. Ready-to-use.

5.7. Substrate

15 mL, TMB (tetramethylbenzidine). Ready-to-use.

5.8. Stop Solution

15 mL, 0.5 M sulfuric acid. Ready-to-use.

5.9. Sample Diluent

60 mL, PBS/BSA buffer. Addition of 0.095 % sodium azide. Ready-to-use.

5.10. Washing Buffer

60 mL, PBS + Tween 20, 10x concentrate. Final concentration: dilute 1+9 with distilled water. If during the cold storage crystals precipitate, the concentrate should be warmed up at 37°C for 15 minutes.

5.11. Plastic Foils

2 pieces to cover the microtiter strips during the incubation.

5.12. Plastic Bag

Resealable, for the dry storage of non-used strips.

6. Materials Required but not Provided

- 5 µL-, 100 µL- and 500 µL micro- and multichannel pipets
- Microtiter Plate Reader (450 nm)
- Microtiter Plate Washer
- Reagent tubes for the serum dilution
- Bidistilled water

7. Specimen Collection and Handling

Principally serum or plasma (EDTA, heparin) can be used for the determination. Serum is separated from the blood, which is aseptically drawn by venipuncture, after clotting and centrifugation. The serum or plasma samples can be stored refrigerated (2-8°C) for up to 48 hours, for a longer storage they should be kept at -20 °C. The samples should not be frozen and thawed repeatedly. Lipemic, hemolytic or bacterially contaminated samples can cause false positive or false negative results.

For the performance of the test the samples (not the standards) have to be diluted 1:101 with ready-to-use sample diluent (e.g. 5 µL serum + 500 µL sample diluent).

8. Assay Procedure

8.1. Preparation of Reagents

Washing Solution: dilute before use 1+9 with distilled water. If during the cold storage crystals precipitate, the concentrate should be warmed up at 37°C for 15 minutes.

- Strict adherence to the protocol is advised for reliable performance. Any changes or modifications are the responsibility of the user.
- All reagents and samples must be brought to room temperature before use, but should not be left at this temperature longer than necessary.
- Standards and samples should be assayed in duplicates.
- A standard curve should be established with each assay.
- Return the unused microtiter strips to the plastic bag and store them dry at 2-8°C.

8.2. Assay Steps

1. Prepare a sufficient amount of microtiter wells for the standards, controls and samples in duplicate as well as for a substrate blank.
2. Pipet 100 µL each of the **diluted** (1:101) samples and the **ready-to-use** standards and controls respectively into the wells. Leave one well empty for the substrate blank.
3. Cover plate with the enclosed foil and incubate at room temperature for 60 minutes.
4. Empty the wells of the plate (dump or aspirate) and add 300 µL of diluted washing solution. This procedure is repeated totally three times. Rests of the washing buffer are afterwards removed by gentle tapping of the microtiter plate on a tissue cloth.
5. Pipet 100 µL each of ready-to-use conjugate into the wells. Leave one well empty for the substrate blank.
6. Cover plate with the enclosed foil and incubate at room temperature for 30 minutes.
7. Empty the wells of the plate (dump or aspirate) and add 300 µL of diluted washing solution. This procedure is repeated totally three times. Rests of the washing buffer are afterwards removed by gentle tapping of the microtiter plate on a tissue cloth.
8. Pipet 100 µL each of the ready-to-use substrate into the wells. This time also the substrate blank is pipetted.
9. Cover plate with the enclosed foil and incubate at room temperature for 20 minutes in the dark (e.g. drawer).
10. To terminate the substrate reaction, pipet 100 µL each of the ready-to-use stop solution into the wells. Pipet also the substrate blank.
11. After thorough mixing and wiping the bottom of the plate, perform the reading of the absorption at 450 nm (optionally reference wavelength of 620 nm). The color is stable for at least 60 minutes.

9. Evaluation

The mean values for the measured absorptions are calculated after subtraction of the substrate blank value. The difference between the single values should not exceed 10%.

Example

	OD Value	corrected OD	Mean OD Value
Substrate Blank	0.018		
Negative Control	0.052 / 0.048	0.034 / 0.030	0.032
Cut-Off Standard	0.513 / 0.521	0.495 / 0.503	0.499
Weak Positive Control	1.317 / 1.300	1.299 / 1.282	1.291
Positive Control	2.177 / 2.153	2.159 / 2.135	2.147

The above table contains only an example, which was achieved under arbitrary temperature and environmental conditions. The described data constitute consequently **no reference values** which have to be found in other laboratories in the same way.

9.1. Qualitative Evaluation

The calculated absorptions for the patient sera, as mentioned above, are compared with the value for the cut-off standard. If the value of the sample is higher, there is a positive result.

For a value below the cut-off standard, there is a negative result. It seems reasonable to define a range of +/-20 % around the value of the cut-off as a grey zone. In such a case the repetition of the test with the same serum or with a new sample of the same patient, taken after 2-4 weeks, is recommended. Both samples should be measured in parallel in the same run.

The positive control must show at least the double absorption compared with the cut-off standard.

9.2. Quantitative Evaluation

The ready-to-use standards and controls of the Mumps antibody kit are defined and expressed in arbitrary units (U/mL). This results in an exact and reproducible quantitative evaluation. Consequently for a given patient follow-up controls become possible. The values for controls and standards in units are printed on the labels of the vials.

For a quantitative evaluation the absorptions of the standards and controls are graphically drawn against their concentrations. From the resulting reference curve the concentration values for each patient sample can then be extracted in relation to their absorptions. It is also possible to use automatic computer programs.

10. Assay Characteristics

Mumps ELISA	IgG	IgA	IgM
Intra-Assay-Precision	8.9 %	8.5 %	8.7 %
Inter-Assay-Precision	7.8 %	7.2 %	6.8 %
Inter-Lot-Precision	1.3 – 13.8 %	1.7 – 9.0 %	4.2 – 9.5 %
Analytical Sensitivity	1.28 U/mL	1.09 U/mL	1.23 U/mL
Recovery	72 – 101 %	104 – 112 %	121 – 128 %
Linearity	82 – 130 %	83 – 124 %	78 – 135 %
Cross-Reactivity	No cross-reactivity to Measles and Varicella zoster		
Interferences	No interferences to bilirubin up to 0.3 mg/mL, hemoglobin up to 8.0 mg/mL and triglycerides up to 5.0 mg/mL		
Clinical Specificity	88 %	100 %	75 %
Clinical Sensitivity	100 %	100 %	100 %

11. References

1. Bayas, JM. et al. Susceptibility to measles, rubella and parotitis in young adults. *Med. Clin. (Barc)*, **106**(15): 561-4 (1996).
2. Chomel, JJ. et al. Rapid direct diagnosis of mumps meningitis by ELISA capture technique. *J. Virol. Methods*, **68**(1): 97-104 (1997).
3. Gut, JP. et al. Symptomatic mumps virus reinfections. *J. Med. Virol.*, **45**(1): 17-23 (1995).
4. Johnson, CE. et al. Antibody persistence after primary measles-mumps-rubella vaccine and response to a second dose given at four to six vs. eleven to thirteen years. *Pediatr. Infect. Dis. J.*, **15**(8): 687-92 (1996).
5. King, SM. et al. Response to measles, mumps and rubella vaccine in paediatric bone marrow transplant recipients. *Bone Marrow Transplant.*, **17**(4): 633-6 (1996).
6. Matter, L. et al. Age-stratified seroprevalence of measles, mumps and rubella (MMR) virus infections in Switzerland after the introduction of MMR mass vaccination. *Eur. J. Epidemiol.*, **13**(1): 61-6 (1997).
7. Pison Garces, FJ. et al. Immunity to measles, mumps and rubella in children vaccinated with triple viral vaccine. *Aten. Primaria*, **15**(4): 235-7 (1995).