

TruQuick™ H. pylori Ab 40T

A rapid test for the qualitative detection of antibodies to *Helicobacter pylori* (*H. pylori*) in whole blood, serum or plasma.

REF TQ5540

IVD

Rx Only

INTENDED USE

TruQuick H. pylori Ab is a rapid immunoassay for the qualitative detection of antibodies to *H. pylori* in whole blood, serum, or plasma to aid in the diagnosis of *H. pylori* infection.

SUMMARY AND EXPLANATION OF THE TEST

H. pylori is a small, spiral-shaped bacterium that lives in the surface of the stomach and duodenum. It is implicated in the etiology of a variety of gastrointestinal diseases, including duodenal and gastric ulcer, non-ulcer dyspepsia and active and chronic gastritis.^{1, 2} Both invasive and non-invasive methods are used to diagnose *H. pylori* infection in patients with symptoms of gastrointestinal disease. Specimen dependent and costly invasive diagnostic methods include gastric or duodenal biopsy followed by urease testing (presumptive), culture, and/or histologic staining.³ Non-invasive techniques include the urea breath test, which requires expensive laboratory equipment and moderate radiation exposure, and serological methods.^{4, 5} Individuals infected with *H. pylori* develop antibodies which correlate strongly with histologically confirmed *H. pylori* infection.⁶⁻⁸ TruQuick H. pylori Ab is a simple test that utilizes a combination of *H. pylori* antigen-coated particles and anti-human IgG to qualitatively and selectively detect *H. pylori* antibodies in whole blood, serum, or plasma.

BIOLOGICAL PRINCIPLES

TruQuick H. pylori Ab is a qualitative membrane based immunoassay for the detection of *H. pylori* antibodies in whole blood, serum, or plasma. In this test procedure, anti-human IgG is immobilized in the test line region of the test. After specimen is added to the specimen well of the device, it reacts with *H. pylori* antigen coated particles in the test. This mixture migrates along the length of the test and interacts with the immobilized anti-human IgG. If the specimen contains *H. pylori* antibodies, a colored line will appear in the test line region indicating a positive result. If the specimen does not contain *H. pylori* antibodies, a colored line will not appear in this region indicating a negative result. To serve as a procedural control, a colored line will always appear in the control line region, indicating that proper volume of specimen has been added and membrane wicking has occurred.

REAGENTS/MATERIALS PROVIDED

The maximum number of tests obtained from this test kit is listed on the outer box.

- Test Cassettes: The test contains *H. pylori* antigen-coated particles and anti-human IgG coated on the membrane. Each cassette is packaged in a foil pouch.
- Buffer: a buffered solution containing Proclin 300 as a preservative. The Buffer is supplied in a dropper vial ready for use.
- Droppers
- Package insert

MATERIALS NOT PROVIDED

- Specimen collection containers
- Lancets (for fingerstick whole blood only)
- Heparinized capillary tubes and dispensing bulb (for fingerstick whole blood only)
- Centrifuge
- Timer

PRECAUTIONS

1. All reagents are for in vitro diagnostic use only. Do not use after the expiration date.
2. Do not eat, drink or smoke in the area where the specimens or kits are handled.
3. Do not use test if pouch is damaged
4. Handle all specimens as if they contain infectious agents. Observe established precautions against microbiological hazards throughout testing and follow the standard procedures for proper disposal of specimens.
5. Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are being tested.
6. The used test should be discarded according to local regulations.
7. Humidity and temperature can adversely affect results.

HAZARD and PRECAUTIONARY STATEMENTS

Refer to the SDS, available at www.meridianbioscience.com for Hazard and Precautionary Statements.

SHELF LIFE AND STORAGE

Store as packaged in the sealed pouch at room temperature or refrigerated (2-30 C). The test is stable through the expiration date printed on the sealed pouch. The test must remain in the sealed pouch until use. **DO NOT FREEZE.** Do not use beyond the expiration date.

SPECIMEN COLLECTION AND PREPARATION


1. TruQuick H. pylori Ab can be performed using whole blood (from venipuncture or fingerstick) serum or plasma.
2. To collect **Fingerstick Whole Blood specimens**:
 - Wash the patient's hand with soap and warm water or clean with an alcohol swab. Allow to dry.
 - Massage the hand without touching the puncture site by rubbing down the hand towards the fingertip of the middle or ring finger.
 - Puncture the skin with a sterile lancet. Wipe away the first sign of blood.
 - Gently rub the hand from wrist to palm to finger to form a rounded drop of blood over the puncture site.
 - Add the Fingerstick Whole Blood specimen to the test by using **a capillary tube**:

- Touch the end of the capillary tube to the blood until filled to approximately 75 µL. Avoid air bubbles.
 - Place the bulb onto the top end of the capillary tube, then squeeze the bulb to dispense the whole blood to the specimen area of the Test Cassette.
- Add the Fingerstick Whole Blood specimen to the test by using **hanging drops**:
- Position the patient's finger so that the drop of blood is just above the specimen area of the Test Cassette.
 - Allow three hanging drops of fingerstick whole blood to fall into the center of the specimen area on the Test Cassette, or move the patient's finger so that the hanging drop touches the center of the specimen area. Avoid touching the finger directly to the specimen area.
3. Separate serum or plasma from blood as soon as possible to avoid hemolysis. Use only clear non-hemolyzed specimens.
 4. Testing should be performed immediately after the specimens have been collected. Do not leave the specimens at room temperature for prolonged periods. Serum and plasma specimens may be stored at 2-8 C for up to three days. For long term storage, specimens should be kept below -20 C. Whole blood collected by venipuncture should be stored at 2-8 C if the test is to be run within two days of collection. Do not freeze whole blood specimens. Whole blood collected by fingerstick should be tested immediately.
 5. Bring specimens to room temperature prior to testing. Frozen specimens must be completely thawed and mixed well prior to testing. Specimens should not be frozen and thawed repeatedly.
 6. If specimens are to be shipped, they should be packed in compliance with local regulations covering the transportation of etiologic agents.

TEST PROCEDURE

Allow the test, specimen, Buffer and/or controls to reach room temperature (15-30 C) prior to testing.

1. Bring the pouch to room temperature before opening it. Remove the Test Cassette from the sealed pouch and use it as soon as possible. Place the cassette on a clean and level surface.
 - For **Serum or Plasma** specimen:
 - Hold the dropper vertically and transfer **3 drops of serum or plasma** (approximately 75 µL) to the specimen well of Test Cassette and start the timer. See illustration below.
 - For **Venipuncture Whole Blood** specimen:
 - Hold the dropper vertically and transfer **3 drops of whole blood** (approximately 75 µL) to the specimen well, then add **1 drop of Buffer** (approximately 40 µL), and start the timer. See illustration below.
- For **Fingerstick Whole Blood** specimen:
 - To use a capillary tube: Fill the capillary tube and transfer **approximately 75 µL of fingerstick whole blood specimen** to the specimen area of Test Cassette, then add **1 drop of Buffer (approximately 40 µL)** and start the timer. See illustration below.
 - To use hanging drops: Allow **3 hanging drops of fingerstick whole blood specimen** (approximately 75 µL) to fall into the specimen area of Test Cassette, then add **1 drop of Buffer (approximately 40 µL)** and start the timer. See illustration below.
3. Wait for the colored line(s) to appear. **Read results at 10 minutes.** Do not interpret the result after 20 minutes.



INTERPRETATION OF RESULTS

(Please refer to the illustration above.)

POSITIVE: Two distinct colored lines appear. One colored line should be in the control region (C) and another colored line should be in the test region (T).

***NOTE:** The intensity of the color in the test line region (T) will vary depending on the concentration of *H. pylori* antibody present in the specimen. Therefore, any shade of color in the test region (T) should be considered positive.

NEGATIVE: One colored line appears in the control region (C). No apparent colored line appears in the test region (T).

INVALID: Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new Test Cassette. If the problem persists, discontinue using the test kit immediately and contact Meridian's Technical Service Department at 800-343-3858 or your local distributor.

QUALITY CONTROL

This test should be performed per applicable local, state, or federal regulations or accrediting agencies.

A procedural control is included in the test. A colored line appearing in the control region (C) is the internal procedural control. It confirms sufficient specimen volume and correct procedural technique.

External controls are not supplied with this kit; however, it is recommended that positive and negative external controls be tested as a good laboratory practice to confirm the test procedure and to verify proper test performance.

If the expected control reactions are not observed, repeat the control tests as the first step in determining the root cause of the failure. If control failures are repeated please contact Meridian's Technical Services Department at 1-800-343-3858 (US) or your local distributor.

EXPECTED VALUES

TruQuick H. pylori Ab has been compared with Culture/Histology, demonstrating an overall accuracy of 94.6%.

LIMITATIONS OF THE PROCEDURE

1. TruQuick H. pylori Ab is for in vitro diagnostic use only. The test should be used for the detection of *H. pylori* antibodies in whole blood, serum or plasma specimens only. Neither the quantitative value nor the rate of increase in *H. pylori* antibody concentration can be determined by this qualitative test.
2. TruQuick H. pylori Ab will only indicate the presence of *H. pylori* antibodies in the specimen and should not be used as the sole criteria for the diagnosis of *H. pylori* infection.
3. As with all diagnostic tests, all results must be interpreted together with other clinical information available to the physician.
4. If the test result is negative and clinical symptoms persist, additional testing using other clinical methods is recommended. A negative result does not at any time preclude the possibility of *H. pylori* infection.

SPECIFIC PERFORMANCE CHARACTERISTICS

Clinical Sensitivity, Specificity and Accuracy

TruQuick H. pylori Ab has been evaluated with specimens obtained from a population of symptomatic and asymptomatic individuals who presented for endoscopic examination. Biopsy (Culture) served as the reference method for TruQuick H. pylori Ab. Histology and a Rapid Urease Test (RUT) were performed on all negative culture specimens. The specimen was considered positive if Culture was positive. The specimen was also considered positive if the Culture was negative, but both Histology and RUT were positive. The result shows that the sensitivity of TruQuick H. pylori Ab is 96.8% and the specificity is 93.0% relative to Biopsy/Histology/RUT.

TruQuick H. pylori Ab vs. Biopsy/Histology/RUT

Method	Biopsy/Histology/RUT		Total Results
	Positive	Negative	
TruQuick H. pylori Ab	150	15	165
	5	200	205
Total Results	155	215	370

Sensitivity: 96.8% (95% CI*: 92.6%-98.9%)

Specificity: 93.0% (95% CI*: 88.8%-96.0%)

Correlation: 94.6% (95% CI*: 91.8%-96.7%)

*Confidence Interval

REPRODUCIBILITY

Intra-Assay Precision

Within-run precision was determined by using 10 replicates of four specimens: a negative, a low positive, a medium positive and a high positive. The negative, low positive, medium positive and high positive values were correctly identified > 99% of the time.

Inter-Assay Precision

Between-run precision was determined by 10 independent assays on the same four specimens: a negative, a low positive, a medium positive and a high positive. Three different lots of TruQuick H. pylori Ab were tested using negative, low positive, medium positive and high positive specimens. The specimens were correctly identified > 99% of the time

CROSSREACTIVITY

Sera containing known amounts of antibodies to *H. pylori* have been tested from patients diagnosed with HAV, HBV, HCV, HEV, HIV, HAMA, RF, Mononucleosis, Toxoplasmosis, CMV and Syphilis. No crossreactivity was observed, indicating that TruQuick H. pylori Ab has a high degree of specificity for antibodies to *H. pylori*.

TESTS FOR INTERFERING SUBSTANCES

The following substances were added to plasma and serum pods containing a portion of a moderate positive sample. None of the following interferents reacted at the concentrations tested.

Ascorbic acid 20 mg/dL	Uric acid 20 mg/dL
Hemoglobin 1 g/dL	Acetaminophen 20 mg/dL
Genistic acid 20 mg/dL	Acetylsalicylic acid 20 mg/dL
Oxalic acid 60 mg/dL	Methanol 10%
Bilirubin 1 g/dL	Creatine 200 mg/dL
	Human albumin 2 mg/dL
	Caffeine 20 mg/dL

REFERENCES

1. Marshall BJ, McGeech DB, Rogers PAR, Glancy, RG. Pyloric Campylobacter infection and gastrooduodenal disease. Med J Australia 1985;149:439-44.
2. Soil AH. Pathogenesis of peptic ulcer and implications for therapy. N Engl J Med 1990;322:909-16.
3. Hazell SL, et al. Campylobacter pyloridis and gastritis I: Detection of urease as a marker of bacterial colonization and gastritis. Am J Gastroenterol 1987;82(4):292-96.
4. Loffeld RJLF, et al. Usefulness of several commercial enzyme-linked immunoassays for detection of Helicobacter pylori infection in clinical medicine. Euro J Gastroen Hepa. 1993;333-37.
5. Cutler AF, et al. Accuracy of invasive and non-invasive tests to diagnose Helicobacter pylori infection. Gastroenterology. 1995;109:136-141.
6. Ansong R, Von Recklinghausen G, Pomarius R, Schmid EN. Evaluation of techniques for isolation, subcultivation and preservation of Helicobacter pylori. J Clin Micro 1991;29:51-53.

7. Pronovost AP, Rose SL, Pawlak J, Robin H, Schneider R. Evaluation of a new immunodiagnostic assay for Helicobacter pylori antibody detection: Correlation with histopathological and microbiological results. J Clin Micro 1994;32:46-50.
8. Megraud F, Bassens-Rabbe MP, Denis F, Belbouri A, Hoa DQ. Seroepidemiology of Campylobacter pylori infection in various populations. J Clin Micro 1989;27:1870-3.



SNTQ5540






















REV. 04/17

 Manufactured By	<p>Meridian Bioscience, Inc. Corporate Office 3471 River Hills Drive Cincinnati, Ohio 45244 USA Telephone: 513.271.3700 Orders/Customer Service: 800.543.1980 Technical Support Center: 800.343.3858 Information Fax: 513.272.5432 Ordering Fax: 513.271.0124</p>
 Authorized Representative	<p>Meridian Bioscience Europe S. r. l. Via dell' Industria, 7 20020 Villa Cortese, Milano ITALY Tel: +39 0331 43 36 36 Fax: +39 0331 43 36 16 Email: info@meridianbioscience.eu WEB: www.meridianbioscience.eu</p>

SYMBOL USAGE

You may see one or more of these symbols on the labeling/packaging of this product:

Key guide to symbols

	Use By	CONTROL +	Positive control
	Batch Code	CONTROL -	Negative control
	In vitro diagnostic medical device	EC REP	Authorized Representative in the European Community
	This product fulfills the requirements of Directive 98/79/EC on in vitro diagnostic medical devices	SMP PREP DIL SPE	Sample Preparation Apparatus containing Sample Diluent
	Catalogue number		Do not freeze
	Consult Instructions for Use	BUF RXN	Reaction Buffer
	Manufacturer		For IVD Performance Evaluation Only
	Contains sufficient for <n> tests	SOLN STOP	Stopping Solution
	Temperature limitation	CONJ ENZ	Enzyme Conjugate
	Serial number	CONTROL	Assay Control
	Test Device	REAG	Reagent
	Date of manufacture	BUF WASH	Wash Buffer
	Buffer		Warning
	Conjugate	DIL SPE	Specimen Diluent (or Sample Diluent)
	Substrate	BUF WASH 20X	Wash Buffer Concentration: 20X
	Research Use Only	DET REAG	Detection Reagent
	Investigational Use Only	R_x Only	Prescription Use Only
	Do not use if package is damaged		

For technical assistance, call Technical Support Services at 800-343-3858 between the hours of 8AM and 6PM, USA Eastern Standard Time. To place an order, call Customer Service Department at 800-543-1980.